



# Clifton Health Department Food Recall Notification



## Food/Drug/Pet Recall: ACON Flowflex SARS-CoV-2 Antigen Rapid Test (Self-Testing)

**Update: The U.S. Food and Drug Administration (FDA) updated this safety communication to clarify that the ACON Flowflex SARS-CoV-2 Antigen Rapid Test (Self-Testing) in dark blue packaging is a product of ACON Biotech (Hangzhou) Co., Ltd.**

The FDA issued a safety communication warning people not to use the ACON Biotech test named Flowflex SARS-CoV-2 Antigen Rapid Test (Self-Testing). This test is packaged in a dark blue box (see image below) and has not been authorized, cleared, or approved by the FDA for distribution or use in the United States. The FDA is concerned about the risk of false results when using this unauthorized test.