



# Clifton Health Department Food Recall Notification



Food/Drug/Pet Recall:

Skippack Medical Lab

SARS-CoV-2 Antigen Rapid Test (Colloidal Gold)

the FDA is warning people to not use the Skippack Medical Lab SARS-CoV-2 Antigen Rapid Test (Colloidal Gold). This test is not authorized, cleared, or approved by the FDA for distribution or use in the United States. The FDA is concerned about the risk of false results when using this test because SML Distribution LLC has not provided the FDA with adequate data to show the tests work correctly.

SML Distribution LLC is recalling the Skippack Medical Lab SARS-CoV-2 Antigen Rapid Tests (Colloidal Gold), and the FDA has identified this issue as a Class I recall, the most serious type of recall.

The FDA has not received reports of injuries or death associated with use of the Skippack Medical Lab SARS-CoV-2 Antigen Rapid Test (Colloidal Gold).